

Produktbeskrivning - Precision balance

ALL MODELS

Model	Weighing capacity [Max] g	Read-ability [d] g	Verific. value [e] g	Minimum load [Min] g	Linearity g	Option Verification	Option DAkkS Calibration Certificate
<u>PNS 600-3</u>	620	0,001			0,004	–	✓
<u>PNS 3000-2</u>	3200	0,01			0,02	–	✓
<u>PNS 12000-1</u>	12000	0,1			0,2	–	✓
Note: For applications that require verification, please order verification at the same time, initial verification at a later date is not possible. Verification at the factory,							
<u>PNJ 600-3M</u>	620	0,001	0,01	0,02	0,004	✓	✓
<u>PNJ 3000-2M</u>	3200	0,01	0,1	0,5	0,02	✓	✓
<u>PNJ 12000-1M</u>	12000	0,1	1	5	0,2	✓	✓